

Calibration Report: Pressure Transducer

25 November 2003

Bryan Fabbri
Analytical Services and Materials, Inc.
Hampton, Virginia

Summary

Calibration Date: 25 November 2003

Next Calibration Date: 25 November 2004

A collection, analysis and calibration of data from Pressure Transducer, S/N 1370787 has been completed. The calibrations were performed by the manufacturer Setra Systems, Inc. These data were collected by Setra Systems, Inc. 25 November 2003.

MODEL : 270

SERIAL NUMBER : 1370787

RANGE: 800–1100 HPA/Mb

The test data presented in data table format show the transducer output in VDC as it is to its applied pressure in millibars (Mb). The Non-linearity errors and extrapolated errors in %FS (full scale) are noted. This unit meets the specifications detailed on the next page.

APPLICATION: Add corrections to measurements per calibration data if applicable.

Setra

Calibration Certificate

Technician: LK *X.K.*
Part No: 270100-01-623
Model: 270

Serial No: 1370787 Range: 800 to 1100 HPA/MB
Work Order: R55542 Nom. Output: 0 to 5 VDC
Date: 11/25/2003 Supply: 12VDC

CALIBRATION DATA

APPLIED PRESSURE (HPA/MB)	TRANSDUCER OUTPUT (VDC)	NONLINEARITY ERRORS (% FS)	EXTRAPOLATED ERRORS (% FS)
800.0000	-0.0003	0.000	Zero -0.005
900.0000	1.6667	0.003	
1000.0000	3.3330	-0.009	Span 0.011
1100.0000	5.0003	0.000	

SPECIFICATIONS

1. Nonlinearity: +/- 0.05 %FS, END POINT method, ISA.#S-37.1
2. Zero pressure output: 0 VDC +/- 0.1 %FS
3. Full Scale output: 5 VDC +/- 0.1 %FS
4. This unit meets the specifications defined above.

NOTES

1. All errors are expressed as: Percent Full-Scale output.
2. Consult specification sheet for additional specifications.
3. This calibration is certified per N.I.S.T. traceable primary standards.
Primary standard: DHI, NIST# 821/265025-01
Primary standard cal. date: 4-30-03
Location of cal.: INPCCTR1
4. This part uses spec. record number: 270100-01-623.15
5. This certificate cannot be reproduced except in full, without the written approval of Setra Systems, Inc.

COMMENTS

AS RECEIVED IN SPEC. RECERTIFIED.

OPTIONS: 623, 703, 810

159 Swanson Road, Boxborough, MA 01719/Telephone 1-800-257-3872, (978) 263-1400