Calibration Report: Absolute Cavity Radiometers S.N. 31041 and 31105

F. M. Denn

Analytical Services & Materials, Inc., Hampton, Virginia

SUMMARY

Calibration date: 2002 September 27. Next calibration due: 2003 September 27.

Calibrations of two Absolute Cavity Radiometers have been completed. The World Radiation References (WRRs) and associated uncertainties with respect to SI units (U95%) are as follows:

Absolute Cavity			
Radiometer	Controller	WRR	U95%
31041	34970A	0.99785	0.34
31105	34970A	1.00357	0.35

Application:

 $I = WRR^*(Io) \pm U95\%$

Where:

I = WRR corrected irradiance, Watt/meter². Io = Irradiance output of the cavity-controller system, Watt/meter². U95% = the 95 % confidence interval.

Calibration certificates from the National Renewable Energy Laboratory located in Golden Colorado are included in this document.

DISCUSSION

Calibration data from Absolute Cavity Radiometers were collected at NREL in September 2002. The serial numbers of these sensors are 31041 and 31105. The calibration standards used were those kept at NREL. These calibration data were analyzed to produce a new World Radiation Reference (WRR) factor and 95-percent uncertainty bound (U95), WRT SI units, for each radiometer. These coefficients are compared to prior calibration results. The instrument setup, data collection, data analysis and uncertainty calculation are as reported in the NPC2001 reference.

CALIBRATION HISTORIES

	Test Cavity Serial Number	Controller	WRR	U95% WRT SI
NPC2002	31041	34970A	0.99785	0.34
NPC2001	31041	34970A	0.99793	0.33
NPC2001	31041	406	0.99830	0.35
IPC-IX (2000	0) 31041	406	0.99799	0.55
NPC1999	31041	406	0.99827	0.39
NPC1998	31041	406	0.99833	0.37
NPC1997	31041	406	0.99961	0.42
NPC2002	31105	34970A	1.00357	0.35
NPC2001	31105	34970A	1.00327	0.34

Calibrations labeled NPCyear took place at the National Renewable Energy Laboratory in Golden, Colorado. Calibrations labeled IPC took place at the World Radiation Center in Davos, Switzerland.

National Renewable Energy Laboratory Solar Radiation Research Laboratory Metrology Laboratory Calibration Certificate for Absolute Cavity Radiometer

NREL Pyrheliomter Comparisons, NPC-2002

 Organization: NASA/LARC-AS&M
 Operator Name: Fred Denn

 Model Number: AHF
 Serial Number: 31041

 Control Unit Serial Number: SG41001207
 Manufacturer Cal. Factor: 1.99992

 Heater Resistance: 155.0 Ω
 Default Sensitivity: 0.0105 µv/W/cm²

 Lead Resistance: 0.066 Ω
 Circuit Resistance: 4.18 Ω

 Thermistor Coefficients: 0.0010295, 0.0002391, and 0.0000001568
 Due Date: 09/27/2003

 Environmental Conditions: (see attached Figures)
 Procedure: NREL/TP-463-20619

Standards Used:

Serial Number	Operator	WRR'	Calibration Date	Due Date
AHF28968	Reda	0.99866	October, 2000 *	October, 2005
AHF29220	Reda	0.99846	October, 2000 *	October, 2005
AHF30713	Reda	0.99861	October, 2000 *	October, 2005
TMI68018	Reda	0.99848	October, 2000*	October, 2005

* Ninth International Pyrheliometer Comparisons (IPC-IX), PMOD, Davos, Switzerland

Results with traceability to the World Radiometric Reference (WRR):

-	WRR Transfer Factor (WRR-TF)	: 0.99785
-	Uncertainty, U ₉₅	: 0.34 %
-	Coverage Factor	:2
	Contrago Antion	ST-70

Data Analysis by	: Ibrahim Reda	QA by	: Tom Stoffel
Signature	: I.R.L	Signature	Tom Stepl
Date	: 10/02/2002	Date	: 10/02/2002

This calibration certificate applies only to the item identified above and shall not be reproduced other than in full, without specific written approval by the calibration facility. Calibration certificates without signatures are not valid.

NREL • 1617 Cole Boulevard • Golden, Colorado 80401-3393 • USA • (303) 275-3000 Operated for the U.S. Department of Energy by Midwest Research Institute • Battelle • Bechtel




Page 1 of 3

Calibration Event	Manufacturer Calibration Factor (MCF)	WRR-TF	Resultant Calibration Factor (RCF)
NPC-1997	Not Provided	0.99961	
NPC-1998	Not Provided	0.99833	
NPC-1999	Not Provided	0.99827	-
NPC-2001	1.99992	0.99793	1.99578
NPC-2002	1.99992	0.99785	1.99562

History for 31041 Calibration

* RCF = MCF * WRR-TF

Reda, 10/2/2002



WRR-Transfer Factor vs Mountain Standard Time for AHF31041 on September 23-27, 2002

National Renewable Energy Laboratory Solar Radiation Research Laboratory Metrology Laboratory

Calibration Certificate for Absolute Cavity Radiometer NREL Pyrheliomter Comparisons, NPC-2002

Organization: NASA/LARC-AS&M	Operator Name: Fred Denn
Model Number: AHF	Serial Number: 31105
Control Unit Serial Number: US37030621	Manufacturer Cal. Factor: 1.9989
Heater Resistance: 155.4 Ω	Default Sensitivity: 0.0105 µv/W/cm2
Lead Resistance: 0.066 Q	Circuit Resistance: 2.55 Q
Thermistor Coefficients: 0.0010295, 0.0002391	l, and 0.0000001568
Calibration Date: 09/27/2002	Due Date: 09/27/2003
Environmental Conditions: (see attached Figu	res)
Procedure: NREL/TP-463-20619	

Standards Used:

Serial Number	Operator	WRR'	Calibration Date	Due Date
AHF28968	Reda	0.99866	October, 2000 *	October, 2005
AHF29220	Reda	0.99846	October, 2000 *	October, 2005
AHF30713	Reda	0.99861	October, 2000 *	October, 2005
TM168018	Reda	0.99848	October, 2000*	October, 2005
	+			

* Ninth International Pyrheliometer Comparisons (IPC-IX), PMOD, Davos, Switzerland

Results with traceability to the World Radiometric Reference (WRR):

-	WRR Transfer Factor	(WRR-TF)	: 1.00357	
---	---------------------	----------	-----------	--

 Uncerta 	unty, Uos	: 0.35 %
-----------------------------	-----------	----------

- Coverage Factor :2

Data Analysis by: Ibrahim RedaQA by: Tom StoffelSignature: I.R.d.SignatureSignatureDate: 10/02/2002Date: 10/02/2002

This calibration certificate applies only to the item identified above and shall not be reproduced other than in full, without specific written approval by the calibration facility. Calibration certificates without signatures are not valid.

NREL • 1617 Cole Boulevard • Golden, Colorado 80401-3393 • USA • (303) 275-3000 Operated for the U.S. Department of Energy by Midwest Research Institute • Battelle • Bechtel




Page 1 of 3

Calibration Event	Manufacturer Calibration Factor (MCF)	WRR-TF	Resultant Calibration Factor (RCF)
NPC-1997	Not Provided	0.99776	
NPC-1999	Not Provided	1.00021	-
NPC-2001	1.9989	1.00327	2.00544
NPC-2002	1.9989	1.00357	2.00604
10			

History for 31105 Calibration

* RCF = MCF * WRR-TF

Reda, 10/2/2002



REFERENCES

Reda, I., Stoffel, T., "Results of NREL Pyrheliometer Comparisons NPC2001", National Renewable Energy Laboratory, Center for Renewable Energy Resources, Measurements & Instrumentation Team, 2001.

Swiss Meteorological Institute, (May 2001). "International Pyrheliometer Comparison IPC-IX." Working Report No. 197, Davos and Zurich.

Reda, I., Stoffel, Wilcox, S., "Results of NREL Pyrheliometer Comparisons NPC1999", National Renewable Energy Laboratory, Center for Renewable Energy Resources, Measurements & Instrumentation Team, 21 September 1999.

Reda, I., Stoffel, T., Treadwell, J., "Results of NREL Pyrheliometer Comparisons NPC1998", National Renewable Energy Laboratory, Center for Renewable Energy Resources, Measurements & Instrumentation Team, 11 November 1998.

Reda, I., Stoffel, T., Treadwell, J., "Results of NREL Pyrheliometer Comparisons NPC1997", National Renewable Energy Laboratory, Center for Renewable Energy Resources, Measurements & Instrumentation Team, 11 November 1997.

Reda, Ibrahim, Stoffel, Tom, "Results of the NREL Pyrheliometer Comparisons NPC1996, 1-5 October 1996", National Renewable Energy Laboratory, Renewable Energy Resources Center, Measurements and Instrumentation Team.