

Calibration Report: Spectroradiometer

s/n: 6207

ECN: 2008990

30 June 2003

Bryan Fabbri
Analytical Services & Materials, Inc.
Hampton, Virginia

SUMMARY

Calibration date: 12 May 2003

Next Calibration due: 12 May 2005

A collection, analysis and calibration of data from Analytical Spectral Devices, Inc. (ASDI), Full Range Fieldspec Radiometer, has been completed. The calibration was performed by the manufacturer, ASDI. These data were collected by ASDI, on 12 May 2003.

Model: FR


Serial Number: 6207

The instrument response files for each foreoptic are installed on the computer operating the instrument. Instrument response files are as follows: **Ni62073.RAW**, **1i62073.RAW**, **5i62073.RAW**, **8i62073.RAW**, **18i62073.RAW**, **COS62073.RAW**.

Irradiance Standard Vendor, Lamp number and File Name, used for Irradiance and Radiance calculations: *Optronic Laboratories, Lamp F627*, **LMP62073.ILL**

Reflectance Standard Vendor, Standard ID, and File Name, used for Radiance calculations: *Labsphere, Target #12137-A*, **BSE62073.REF**

Application: The instrument response files are utilized by the computer operating the instrument at the time of data collection.



ANALYTICAL SPECTRAL DEVICES, INC.

Portable, Accurate Solutions From Light

5335 Sterling Dr., Suite A, Boulder, CO 80301
(303) 444-6522 Fax: (303) 444-6825

Website: www.asdi.com

SPECTRORADIOMETER CERTIFICATE OF CALIBRATION

UNIT AND CALIBRATION NO.: *FSFR 6207/3*

CALIBRATION DATE: *5/12/03* LAMP NO.: *F626* PANEL NO.: *12137-A*

Applicable entries:

Wavelength

Radiometric

Foreoptics:

Bare Fiber

3 Degree

10 Degree

Jumper

1 Degree

5 Degree

18 Degree

Jumper with RCR

2 Degree

8 Degree

RCR

All calibrations have been performed according to Analytical Spectral Devices' accepted procedures, using verifiable NIST-traceable irradiance, reflectance and wavelength standards.

Calibration data resides on the ASD instrument's controlling computer's hard drive and/or the controlling software system disk. Instrument response files: *Ni62073.RAW*, *5i62073.RAW*, *8i62073.RAW*, *18i62073.RAW* and *Cos62073.RAW*.

Irradiance Standard Vendor, Lamp number and File Name, used for Irradiance and Radiance calculations: *Optronic Laboratories, Lamp F626, LMP62073.ILL*

Reflectance Standard Vendor, Standard ID, and File Name, used for Radiance calculations: *Labsphere, Target #12137-A, BSE62073.REF*

ASD Certified Calibration Engineer/Technician:


Signature

Margaret Chisholm


Date


5/30/03

Optronics Laboratory 1000 Watt Irradiance Standard, FEL-626



Spectralon(tm) Reflectance Standard, Target #12137-A






Calibration Accuracy Calculations


λ nm	NIST	Optronics Transfer	.1% Current Error	2mm Position Error	λ cal.	RSS Non Inst.	FR	Irr. RSS	Labsph. Refl.	Rad. RSS
350	1.35%	0.70%	0.75%	0.80%	2.50%	2.04%	2%	2.85%	0.50%	2.90%
654.6	1.01%	0.40%	0.40%	0.80%	0.88%	1.62%	2%	2.57%	0.50%	2.62%
900	1.34%	0.50%	0.35%	0.80%	0.20%	1.86%	2%	2.73%	0.50%	2.77%
1300	1.42%	0.50%	0.30%	0.80%	0.04%	1.91%	2%	2.76%	0.50%	2.81%
1600	1.89%	0.50%	0.20%	0.80%	0.04%	2.27%	2%	3.02%	0.50%	3.06%
2000	3.29%	0.75%	0.14%	0.80%	0.04%	3.56%	2%	4.08%	0.50%	4.11%
2400	6.51%	1.00%	0.14%	0.80%	0.04%	6.68%	2%	6.98%	2.00%	7.26%
2500	6.50%	1.00%	0.14%	0.80%	0.04%	6.67%	2%	6.97%	2.00%	7.25%

NEEL Unit 6207/3; 05/30/03


W/cm²/sr/nm




Mercury Argon Unit 6207/3; 05/12/03
Raw DN




Kaolinite Unit 6207/3; 05/12/03
Reflectance



Mylar Unit 6207/3; 05/12/03
Transmittance





ANALYTICAL SPECTRAL DEVICES, INC.
Portable, Accurate Solutions From Light

5335 Sterling Dr., Suite A, Boulder, CO 80301
(303) 444-6522 Fax: (303) 444-6825

Website: www.asdi.com

FieldSpec Full Range

Quality Assurance Certificate

This document certifies that the instrument listed meets Analytical Spectral Devices, Inc. standard of quality. The instrument has passed and completed all tests and procedures listed and has been found to meet or exceed the following specifications.

Spectral Range: 350-2500nm
Spectral Resolution: 3nm at 700nm
10nm at 1400nm and 2100nm
Scanning Time: 100ms
Sampling Interval: 1.4nm at 350-1050nm
1nm at 1000-2500nm
Wavelength Accuracy: ± 1 nm
Noise Equivalent Radiance: $1.4E-9$ W/cm²/nm/sr at 700nm
 $2.4E-9$ W/cm²/nm/sr at 1400nm
 $8.8E-9$ W/cm²/nm/sr at 2100nm

Comments:

Date: 5/12/2003

Quality Assurance Signature:

Margaret C. [Signature]

Instrument Information

Unit #: 6207 Cal #: 3
SO #: C993740A RMA #: 1142
PC Make: NEC
SN #: 74009951
Operating System: DrDOS

Computer Boot & Setup

AC Power up sequence
 Processor speed
NA Ram
NA Hard Drive
NA Display Settings Cleared
 Parallel Port Mode: Bi-directional
 Power Savings Turned Off
NA Install ViewSpec Pro supplied diskettes
 Install RS^2
NA Black and white option
NA Install Technical Guides

FieldSpec Power

Power on
NA Battery status light
NA Battery power

Software and System Operation

Optimization
 Vnir Noise level
 Swir Noise level
 White reference
 Masked pixels
 Shutter operation
 NEdL
 Radiometric tests - VS Pro & Realtime
NA CD/Floppy drive test - transferring files
 Spectrum save features - multiple
 ForeOptic check

Wavelength Checks

Mylar transmittance, 14 mil thicknes
 Kaolinite sample reflectance
 HgAr in DC corrected Raw Digital Numbers

Configuration File - ASD.INI

Calibration number
 Start/Step
 Masked pixels listing
 Foreoptics listing

Shipping

QA plots: Mylar, Kaolinite, HgAr, NEdL
 Final cal data re-installed and backed up
 RS^2 Software Package Ver. 2.16
 ViewSpec Pro Software Package Ver. 3.09
N/A Special Release Notes, if applicable
 Radiometric Calibration Certificate
lamp #: F627 panel #: 12137-A
 Packing list printed
 Packing list matches Sales Order
 Label computer, disks, p-grip